

# COMPATIBILITY STATEMENT

HTL-STREFA S.A.  
is an importer  
and/or distributor  
of blood lancets  
for single use,  
model BYY3  
(type 561)

This statement is based on successfully passed tests conducted with the lancing devices listed in Appendix 1.

Blood lancets for single use, model BYY3 (type 561), were tested to evaluate their compatibility with certain commercially available lancing devices from various manufacturers.

A non-exhaustive list of compatible lancing devices can be found in Appendix 1.


The brand names of pen injector devices are trademarks of their respective owners and may vary across different markets.

## Appendix 1

List of lancing devices compatible with blood lancets for single use, model BYY3.

**Table 1.** Summary of test results for the compatibility of blood lancets, model BYY3 with the lancing devices.

Manufacturer / Distributor	Brand name	Product picture	Test result Assessment whether the lancing device works with blood lancets for single use (YES/NO)
HTL-STREFA S.A.	droplet® (type 710)		YES
	droplet® (type 700)		YES
A. Menarini	Glucject® Dual		YES
	Glucject® Dual Plus		YES
	MenaLancet® DUAL		YES
Owen Mumford	Autolet® Impression AT271		YES
Abbott	FreeStyle		YES
ADVOCATE	ADVOCATE®		YES
ReliOn	ReliOn™ Lancing Device		YES
Sanofi	BG Star®		YES

Manufacturer / Distributor	Brand name	Product picture	Test result Assessment whether the lancing device works with blood lancets for single use (YES/NO)
Simple Diagnostics	Simple Diagnostics®		YES
Ypsomed	my life Sofflance		YES
ADC	Microlet® next		YES
Bayer	Microlet® 2		YES
PIKDARE S.p.A	Digitest		YES
STERILANCE MEDICAL (SUZHOU) INC.	Digitest (model LDE4)		YES
ReliaMed®	Mini lancing device		YES
Future Diagnostics LLC	CareTouch™		YES
Lifescan Johnson & Johnson	ONETOUCH Delica™		YES